



Cordless endodontic treatment motorized handpiece

EndoSync™ PLUS

INSTRUCTIONS FOR USE



Thank you for purchasing the EndoSync PLUS.

For optimum safety and performance, read this manual thoroughly before using the unit and pay close attention to warnings and notes. Keep this manual in a readily accessible place for quick and easy reference.

Table of Contents

Accident Prevention	3
Precautions	5
Features	6
Parts Identification and Accessories	7
Usage	8
(1) Operating Environments.....	8
(2) Before Use.....	8
Charge Battery.....	8
Connect Contra Angle.....	9
File Installation.....	10
Calibration.....	10
Check Operation.....	11
(3) Operation.....	12
Basic Operation.....	12
Memory Settings.....	13
Memory Setting: OTR Mode.....	14
Memory Setting: Normal Mode (CW).....	15
Memory Setting: Normal Mode (CCW).....	16
Memory Settings: Other Settings.....	17
Restore Default Memories.....	18
(4) After Use.....	18
Take Out File.....	18
Usage; Operation with EndoSync A.I.	19
Install File Electrode.....	19
Connect Transmission Cable.....	20
Check Operation.....	20
Meter Display.....	21
Operation.....	21
Rotor Axle and Built-in Electrode Cleaning.....	23
Maintenance	25
Autoclavable Components.....	25
Wipe with Ethanol (70 vol% to 80 vol%).....	28
Wash and Wipe with Ethanol (70 vol% to 80 vol%).....	29

Replacement Parts and Storage	30
(1) Replacement Parts.....	30
Battery Replacement.....	30
Built-in Electrode Replacement.....	31
(2) Transport and Storage Environments.....	33
Inspection and Warranty	34
Regular Inspection.....	34
Inspection Items.....	34
Warranty.....	35
Troubleshooting	36
Technical Specifications	38
Technical Specifications.....	38
Symbols.....	39
Electromagnetic Disturbances (EMD)	40

Trademarks (™) and Registered Trademarks (®) :

The names of companies, services, etc. used in this manual are either trademarks or registered trademarks.

Accident Prevention

Customers

Make sure to obtain clear instructions concerning the various ways to use this device described in this accompanying manual.

Preventing Accidents

Most operation and maintenance problems result from insufficient attention to basic safety precautions and not being able to foresee potential accidents.

Problems and accidents are best avoided by anticipating potential dangers and operating the device in accordance with the manufacturer's recommendations. First, thoroughly read all precautions and instructions pertaining to safety and accident prevention. Then operate the device with the utmost caution to prevent either damaging the device itself or causing bodily injury.

The following symbols and expressions indicate the degree of danger and harm that could result from ignoring the corresponding instructions:

 **WARNING** This alerts the user of the possibility of extremely serious injury or complete destruction of the device, as well as other property damage including the possibility of fire.

 **CAUTION** This alerts the user of the possibility of minor or moderate injury or damage to the device.

The warning symbols () and caution symbols () that appear next to the main text on the right hand side of the page refer to and are explained by the Warnings and Cautions at the bottom of the page.



This refers to mandatory actions and procedures that are required to prevent damage to the device.

The user (e.g., healthcare facility, clinic, hospital etc.) is responsible for the management, maintenance, and use of medical devices.

This equipment must only be used by dentists and other legally licensed professionals.

Do not use the EndoSync PLUS for anything other than its specified dental purpose.

Rx Only

CAUTION: Federal law restricts this device to sale by or on the order of a dentist. (Valid only for U.S.A.)

Disclaimers

- Brasseler USA will not be responsible for accidents, equipment damage, or bodily injury resulting from:
 1. Repairs made by personnel not authorized by Brasseler USA.
 2. Any changes, modifications, or alterations of its products
 3. The use of products made by other manufacturers, except for those procured by Brasseler USA.
 4. Maintenance or repairs using parts or components other than those specified by Brasseler USA or other than in their original condition
 5. Operating the device in a manner other than described in the operating procedures in this manual or in a manner inconsistent with the safety precautions and warnings in this manual.
 6. Workplace, environmental, or installation conditions that do not conform to those stated in this manual, such as an improper electrical power supply.
 7. Fires, earthquakes, floods, lightning, natural disasters, or forces majeure.
- The useful life of the EndoSync PLUS is 6 years from the date of installation provided it is regularly and properly inspected and maintained.

In Case of Accident

If an accident occurs, the EndoSync PLUS must not be used until repairs have been completed by a qualified and trained technician authorized by the manufacturer.

User Qualifications

Intended Operator Profile

- a) Qualification : Legally qualified person such as dentists for endodontic device operation (it may differs among countries).
 - b) Education and Knowledge : It is assumed that the understands the risks of root canal measuring and treatment. It is also assumed the user is thoroughly familiar with root canal measuring and treatment including the prevention of cross contamination.
 - c) Language Understanding : English (Intended for professional use as described above)
 - d) Experience : Experienced person with operating endodontic device.
- No special training is required except in cases where this is required by legal regulations of the relevant country or region.

Patient Population

Age : Child to Elderly

Weight : Not applicable

Nationality : Not applicable

Sex : Not applicable

Health : It is not intended for use on patients wearing pacemakers or ICDs.

Condition : Conscious and mentally alert person. (Person who can stay still during treatment.)



CAUTION

- When the EndoSync PLUS is connected to the EndoSync A.I., an apex locator, it is not recommended that this device be used with children under 12 years of age.

Precautions

WARNING

- Except for ways described in this manual, this unit must not be connected to or used in combination with any other apparatus or system. It must not be used as an integral component of any other apparatus or system. Brasseler USA will not be responsible for accidents, equipment damage, bodily injury or any other trouble which results from ignoring this prohibition.
- A rubber dam should be used when performing endodontic treatment.
- No modification of this equipment is allowed.
- Do not perform maintenance while using the unit for treatment.
- Instruments which produce considerable electrical noise such as electric scalpels can cause the EndoSync PLUS to operate abnormally. Turn the EndoSync PLUS off before using any instruments that produce electrical noise.
- Do not use this unit on patients who have a pacemaker or an Implantable Cardioverter Defibrillator (ICD). It could cause the pacemaker or the Implantable Cardioverter Defibrillator (ICD) to function abnormally.
- Illumination devices such as fluorescent lights and film viewers which use an inverter can cause the EndoSync PLUS to operate erratically. Do not use the EndoSync PLUS near lights such as these.
- Do not use this unit in the medical operation room.
- Blocked canals cannot be accurately measured.

* Brasseler USA is not responsible for any accidents or other types of trouble that are caused by not following the warnings noted above.

Features

Instructions for how to use the EndoSync PLUS when it is connected to the EndoSync A.I. are printed on a blue background.

■ Liquid Crystal Display (LCD)

The LCD is easy to read and shows all settings as well as how the motor is running.

■ Controls

< OTR (Optimum Torque Reverse) Mode >

If the file torque is less than the set value, the file will keep rotating in the cutting direction.

When the file torque is more than the set value, the file will automatically rotate in the non-cutting and cutting directions repeatedly at the set angle. OTR Mode can set various motor controls as described below. Furthermore, OTR Mode can use files for which the cutting direction is counterclockwise as well as normal files for which the cutting direction is clockwise. (The cutting direction of the file can be set.)



Speed	100, 300, 500, 800, and 1000 rpm.
Reverse Torque Setting	0.2, 0.4, 0.6, 0.8, and 1.0 Ncm (If the rotation speed is 100, 300, or 500 rpm, this can be set to 0.2 or 0.4 Ncm.)
Cutting Direction of File	Clockwise (CW), Counterclockwise (CCW) (The cutting direction is the file rotation direction as seen from the File Release Button side.)
Cutting Angle	90, 150, 180, 270, 360 degrees (If the speed is 100 rpm, this can be set to 90 degrees.)
Non-cutting Angle	30, 60, 90, and 120 degrees
Auto Start & Stop**	The file starts when it is inserted in the canal and stops when it is taken out.
Apical Reverse or Stop**	The motor reverses or stops when the tip of the file reaches a preset position inside the canal.
Reaches Apex**	The motor stops when the tip of the file reaches the apex.

** These functions are available only when the EndoSync PLUS is connected to the EndoSync A.I.

< Normal Mode >

If the file torque is less than the set value, the file will keep rotating in the clockwise direction. When the file torque is more than the set value, the file will automatically start rotating in the counterclockwise direction.

Furthermore, Normal Mode can set various motor controls as described below. Note that files for which the cutting direction is counterclockwise cannot be used for Normal Mode.

Speed	50, 100, 150, 200, 250, 300, 400, 500, 600, 800, and 1000 rpm
Torque Reverse	The motor automatically reverses its rotation if the torque load exceeds the set value to reduce the risk of jamming. (Only when the cutting direction is set to "CW" [clockwise].)
Reverse Torque Settings	0.2, 0.4, 0.6, 0.8, 1.0, 1.5, 2.0, 2.5, 3.0, 4.0, and 5.0 Ncm
Torque Slow Down	The file slows down as torque increases.
Apical Slow Down	The file slows down as it approaches the apex if the EndoSync PLUS is connected to the EndoSync A.I.
Rotation Direction	Clockwise (CW), or Counterclockwise (CCW)
Apical Reverse or Stop**	The motor reverses or stops when the tip of the file reaches a preset position inside the canal.
Auto Start & Stop**	The file starts when it is inserted in the canal and stops when it is taken out.
Apical Torque Reduction**	The automatic torque reverse value is reduced as the file tip approaches the apex.
Reaches Apex**	The motor stops when the tip of the file reaches the apex.

** These functions are available only when the EndoSync PLUS is connected to the EndoSync A.I.

■ Memory

Six combinations of speed, torque etc. can be memorized.

WARNING

- Files for which the cutting direction is counterclockwise can only be used in OTR Mode with CCW setting. These files cannot be used for Normal Mode.
- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".

Parts Identification and Accessories

Parts Identification



Accessories

Standard Accessories

Power Supply Cord (1)	EndoSync Oil (1)	Guide Bar (1)
		<p>* Use the guide bar when cleaning and replacing the built-in electrode.</p> <p> p. 23 "Rotor Axle and Built-in Electrode Cleaning"</p> <p> p. 31 "Built-in Electrode Replacement"</p> 

Optional Accessories

Handpiece Rest (1)	Transmission Cable (1)
	

Usage

(1) Operating Environments

Temperature: +10°C to +40°C (+50°F to +104°F)

Humidity: 30% to 80% (without condensation)

Atmospheric Pressure: 70 kPa to 106 kPa

* If the unit has not been used for some time, make sure it works properly before using it again.

(2) Before Use

Check the following item before using the unit.


- Have autoclavable components been sterilized?  p. 25 “Autoclavable Components”

Charge Battery

The battery is built into the motor handpiece.

* Ambient (room) temperature for charging is from +10°C to +40°C (+50°F to +104°F).

1




Green LED

Connect the power cord to the charger and then plug it in. Turn the charger on. The green Power LED will light up.

 Do not use any charger; use only the one that comes with the EndoSync PLUS.

2



Green LED Orange LED

Put the motor handpiece into the charger as far as it will go. The orange Charge LED will light up to show that charging has begun.

* Charging time is about 120 minutes.

-  Charge the battery as soon as the battery power indicator gets down to its last bar.
-  If the orange charge LED goes off immediately or doesn't light up when the motor handpiece is put into the charger, the battery is probably fully charged. To make sure, take the motor handpiece out and put it back in again.
-  Make sure the contact areas for the motor handpiece and charger are free of debris, especially metal fragments. Wipe with ethanol (70 vol% to 80 vol%) to remove any foreign debris. Do not press down too hard to wipe the charging areas; this could bend the electrical contacts.
-  Do not leave the charger where it will be exposed to direct sunlight.

WARNING

- If an electrical storm occurs while the battery is being charged, do not touch the charger or its cord as there would be a risk receiving an electric shock.
- Do not get the charger wet or use it where it might get wet.

CAUTION

- The battery is not charged when the unit is shipped and must be charged before using the unit.
- Do not pull or yank the cord when disconnecting the power supply cord. Always grip the connectors.
- Use only the power cord provided and plug both ends all the way in.
- Charger and power supply cord must be located outside the so called patient environment (1.5 m around the patient location).

Charge Battery



The number of bars shows how much battery power is left. Recharge the battery when there is only one bar left.





If the battery runs almost completely out, the EndoSync PLUS will automatically turn itself off after about 10 seconds. Recharge the battery as soon as possible.



If the battery power is very low and a large load is applied to the file, the motor may stop or the unit may turn itself off.

This is for safety; there may not be enough power to run the motor with sufficient stability. Recharge the battery if the display shown to the left appears frequently.

Connect Contra Angle



The rotation range of the contra angle is about 290°. Set it so that it reaches the treatment area and the display is easy to see.

 Do not force the contra angle to rotate beyond its limit.

Line up the match marks and push the contra angle onto the motor handpiece until there is an audible click.

* The contra angle must be lubricated with the EndoSync Oil before using for the first time.  p. 27 “Lubrication”




WARNING

- Make sure the connection components for both the motor handpiece and the contra angle are not damaged. An improper connection could cause the motor to reverse unexpectedly and result in injuring the patient.

CAUTION

- Push the contra angle all the way onto the motor handpiece and then give it a light tug to make sure it is securely attached.

File Installation



1. Hold down the file release button.
2. Insert the file and turn it back and forth until it lines up with the latch mechanism.
3. Push the file all the way into the latch. Release the file release button.



- ! It's easiest to hold the button down if you put your index finger where the head joints the body.
- ! Use either Nickel-titanium or stainless steel files.

Calibration

* Before using right after purchase, whenever the motor handpiece or contra angle has been replaced, or if the motor alternates between forward and reverse rotation outside the canal, calibrate the unit in the following way:



1. Turn the unit on. (Press the Main Switch.)
2. Make sure the battery is fully charged (three bars are displayed).
3. Put a commonly used file into the contra angle.
4. Press the Plus or Minus Switch and then select memory M6.
5. Hold down the Select Switch for 2 seconds or more.
6. Set Rotation Mode to "CCW"; see p. 16 "Memory Setting: Normal Mode (CCW)"
7. Turn the unit off. (Hold the Select Switch and then press the Main Switch.)
8. Hold the Plus and Minus Switches and then press the Main Switch to turn the unit back on.
9. When "CO-Adjst" appears in the display, press the Select Switch.
The motor will start running. Make sure there is no load on the file.
10. When "Finished" appears in the display, the motor will stop and the calibration will be completed. Press the Main Switch to go to the Standby Display.

* Once calibration has been completed, you may change the M6 setting and you may turn the unit off with a setting other than M6.

* If the unit uses a file electrode, calibrate the unit by putting the electrode on the file and connecting the unit to a EndoSync A.I. unit which is turned on.


⚠ WARNING

- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".
- Never use deformed or damaged files.
- Give the file a light tug to confirm it is securely held in place. If the file is not securely placed, it could come out and injure the patient.
- When using the linkage function with EndoSync A.I., do not use the files that have no electrical conductivity between the file and its shank. See p. 20 "Check Operation"

⚠ CAUTION

- Use caution when inserting and removing files to avoid injury to fingers.
- Inserting and removing files without holding the file release button down will damage the chuck.
- Make sure the EndoSync PLUS is turned off before inserting or removing files.

Check Operation



Error : 00
Chk - 0M

If a malfunction occurs, the EndoSync PLUS will stop working.
☞ p. 36 "Troubleshooting"

If this error display appears frequently, stop using the unit and contact Brasseler USA. The number that appears after Error will depend on the type of malfunction.

- Make sure the contra angle and motor handpiece are properly and securely connected.
- Make sure the file is securely installed; give it a light tug.
- Check switch operation.

Turn the Main Switch on and use the Plus or Minus Switches to select a memory number. Then press the Main Switch again to see if the EndoSync PLUS runs smoothly.



How to check the EndoSync PLUS's operation when it is connected to the EndoSync A.I.; ☞ pp. 19 to 24 "Usage; Operation with EndoSync A.I."

⚠ WARNING


- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".
- Operate the EndoSync PLUS outside the oral cavity to make sure it will operate properly before using it for treatment.
- Some canals may be impossible to enlarge; always take an X-ray to check.
- Nickel-titanium file may suddenly snap depending on the curvature and shape of the canal; stop using the file if you notice or feel anything amiss.
- Files will eventually break due to metal fatigue and should be replaced before they reach this point.
- Electric noise or a malfunction could interfere with the motor control. Do not depend entirely on the unit controlling itself; always watch the display and be aware of tactile feedback.
- Files will jam and break if too much force is applied to them.
- Files may break even when the torque reverse is turned on, depending on the setting value. Never exert excessive force on the file.
- Files designed for use with engines break easily if too much force is applied. Also do not use these files for canals with excessive curvature.
- Always examine files for stretching and other deformities or damage before using them. Any type of deformity could result in the file breaking.
- Do not let the file release button on the contra angle press against the teeth opposite to the treatment area; this could cause the file to come out and result in an injury.
- Do not press the file release button while the motor is running. It could heat up and cause a burn, or the file could come out and cause an injury.

⚠ CAUTION

- Stop using the EndoSync PLUS if you feel or notice anything unusual. The EndoSync PLUS cannot be used for every canal and should be used along with manual enlargement.
- File break more easily at fast speeds; always follow the file manufacturer's usage recommendations. Also always check the speed settings before use.
- Do not use any type of files except nickel-titanium and stainless steel ones.
- Nickel-Titanium files are easily broken; note the following points.
 - Open the canal up to the apical constriction manually before using a nickel-titanium file.
 - Never use excessive force to insert the file.
 - First remove all foreign matter, such as bits of cotton from the root canal.
 - Never use excessive force to advance the file down the root canal.
 - Do not use for extremely curved canals.
 - Try not to trigger the auto torque reverse function when advancing the file down the canal.
 - Do not skip file sizes; suddenly using a much larger file could break it.
 - If you encounter resistance or the auto torque reverse is triggered, back the file up 3 or 4 mm and carefully advance it down the root canal again. Or replace the file with a smaller size. Never use excessive force.
 - Do not force the file down the root canal or press it against the root canal wall.
 - Do not use the same file continuously in one position as this may create "steps" on the root canal wall.
- Always take file out of the contra angle after use.

(3) Operation

Basic Operation



1. Turn EndoSonic PLUS On: Press the Main Switch.

The Standby Display will appear.

2. Select Memory Number: Press the Plus or Minus Switch.

- * There are six memories for various combinations of speed, torque reverse and rotation mode settings.
- * The backlight will temporarily change color if changing the memory number changes anything other than the speed, torque reverse, and rotation mode settings.

3. Start Motor: Press the Main Switch again.

The Torque Display will appear.

- * If you hold the Main Switch down when you start the motor, it will run only while the switch is held down and stop when the switch is released.
- * You can temporarily change the reverse torque setting while the motor is running by pressing the Plus or Minus Switch. (This works only for Normal Mode.)

! When the Apical Torque Reduction is set to "ON", the torque setting cannot be temporarily changed.

- * The color of the backlight changes based on the load applied to the file.
- * The backlight starts blinking when the load approaches the reverse torque setting. While the OTR is triggered, the backlight does not blink.

4. Stop Motor: Press the Main Switch again.

The Standby Display will reappear.

5. Turn EndoSonic PLUS Off: Hold the Select Switch and then press the Main Switch.

The EndoSonic PLUS turns itself off automatically if it is not used for 3 minutes (initial setting).

For meter readings and operation connected to the EndoSonic A.I.; pp. 21 to 22.

! WARNING

- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".
- Be sure to check the new settings whenever you change the Memory Number.

! CAUTION

- The temperature of the motor handpiece rises up to +47.5°C (+117.5°F) when the ambient temperature is +40°C (+104°F).
- When the OTR seems to be triggered too frequently, or it is triggered immediately after starting the normal rotation, increase the torque setting by one level.

Memory Settings

The initial settings are shown below. These settings can be changed.

Setting	Memory					
	M1	M2	M3	M4	M5	M6
Rotation Mode	OTR	OTR	OTR	OTR	OTR	CW
Cutting Direction of File	CW	CW	CCW	CW	CW	CW
Speed (rpm)	500	500	500	1000	300	500
Reverse Torque (Ncm)	0.6	0.2	0.2	1.0	0.6	1.0
Cutting Angle (degree)	180	180	180	180	180	N/A
Non-cutting Angle (degree)	60	60	60	90	90	N/A
Apical Action Function**	ON	ON	ON	ON	ON	ON
Apical Reverse or Stop**	Stop	Stop	Stop	Stop	Stop	Stop
Auto Start or Stop**	ON	ON	ON	ON	ON	ON
Apical Slow Down**	N/A	N/A	N/A	N/A	N/A	OFF
Apical Torque Reduction**	N/A	N/A	N/A	N/A	N/A	OFF
Torque Slow Down	N/A	N/A	N/A	N/A	N/A	OFF



** These functions are available only when the EndoSync PLUS is connected to the EndoSync A.I.

Operation Mode

There are three modes for canal shaping.

OTR : Used for canal shaping.

CW : Normal 360° forward rotation. Torque reverse and other functions can be used.

CCW : Reverse rotation only.

* When this mode is being used, a double-beep sounds continuously.




WARNING

- Files for which the cutting direction is counterclockwise can only be used in OTR Mode with CCW setting. These files cannot be used for Normal Mode.
- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".

CAUTION

- If the torque limit is too high, the file could jam inside the canal and break.
- The torque settings must be changed depending on the root canal condition and the file.
- If the torque reverse seems to be activated too frequently increase its value.
- When OTR Mode (motor runs back and forth continuously) seems to be triggered too frequently, or it is triggered immediately after starting the normal rotation, increase the torque setting by one line.


Memory Setting: OTR Mode




1. Select a memory number for the Standby Display; press the Plus or Minus Switch.
 2. Hold down the Select Switch for at least 2 seconds to show the display for operation settings.
 3. Press the Select Switch to go from one display to the next.
 4. Press the Plus or Minus Switch to change the setting.
- * Pressing the Main Switch to go back to the Standby Display.
- * The display will go back to the Standby Display when 10 seconds (initial setting) elapses without a switch being pressed.

** These functions are available only when the EndoSync PLUS is connected to the EndoSync A.I.

Memory Setting: Normal Mode (CW)



Memory Setting: Normal Mode (CCW)



1. Select a memory number for the Standby Display; press the Plus or Minus Switch.
2. Hold down the Select Switch for at least 2 seconds to show the display for operation settings.
3. Press the Select Switch to go from one display to the next.
4. Press the Plus or Minus Switch to change the setting.
 - * Pressing the Main Switch to go back to the Standby Display.
 - * The display will go back to the Standby Display when 10 seconds (initial setting) elapses without a switch being pressed.


** These functions are available only when the EndoSync PLUS is connected to the EndoSync A.I.

Memory Settings: Other Settings

Other Settings: The initial settings are shown below.

Beeper (BEEP VOLUME)	High	Backlight (B.L.COLOR CHANGE)	ON
Auto Power Off (AUTO PWR)	3 min.	Return to Standby Time (S.S.R TIME)	10 sec.
Positive/Negative Display (DISP. TYPE)	Posi	Memory Setting Lock (SETTING LOCK)	OFF
Right or Left Handed (DOMI. HAND)	Right		

1. With unit turned off, hold down Select Switch and then press the Main Switch.
2. Press the Select Switch to select one of the settings.
3. Press the Plus or Minus Switches to change the setting.
4. Press the Main Switch to return to the Standby Display.



Restore Default Memories

Restore the initial settings for the memories in the following way.


* This will restore the original memory settings. You cannot restore settings for just one memory.



1. With the unit turned off, hold down the Select Switch, the Plus Switch and the Minus Switch and then turn the unit on with the Main Switch.
2. The "MemClear" display will appear. Press the Select Switch to restore the default memories or press the Main Switch to cancel the operation.
3. Wait until the "Finished" display appears and then press the Main Switch to go to the Standby Display.

(4) After Use

Take Out File



1. Hold down the Select Switch and press the Main Switch to turn the power off.
* The power will go off automatically if the unit is not used and no switches are pressed for 3 minutes.
2. Hold down the file release button and pull the file straight out.




CAUTION

- Take care not to injure your fingers when inserting and removing files.
- Never insert or remove files without holding down the button; this will damage the chuck.
- Make sure the unit is turned off before inserting or removing files.

Usage; Operation with EndoSync A.I.



Install File Electrode



1. Hold down the file release button.
2. Insert the file and turn it back and forth until it lines up with the latch mechanism.
3. Push the file all the way into the latch. Release the file release button.

- ! It's easiest to hold the button down if you put your index finger where the head joints the body.
- ! Use either Nickel-titanium or stainless steel files.

WARNING



- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".
- Some files cannot use the built-in electrode to make measurement; always check for conductivity before using a file.
- Make sure the file goes all the way in. Give it a light tug to make sure it is held securely.
- Never use stretched, deformed or damaged files.
- Make sure the screw is tight enough. Otherwise, it might come out and be swallowed. Also, measurements might not be accurate.

CAUTION

- Never put file in or take them out without pressing the button down. This could damage the chuck. Always hold the button down to put a file in or take it out.
- Use only Ni-Ti or properly designed stainless steel files.
- Be careful not cut your finger when putting files in and taking them out.
- Do not let the cutting part of the file touch the electrode; this will wear it out very quickly.
- Some files cannot be used with this electrode.
- Do not use files shanks larger than the ISO standard: Diameter 2.334 to 2.350 mm
- After use, be sure to take the file out.



Connect Transmission Cable

* Refer to the user manual for the EndoSync A.I.



- ! Do not mix up the cable plugs.
- ! Do not put stress on the transmission cable by twisting, bending, or stretching by wrapping it around the EndoSync PLUS or EndoSync A.I.

Check Operation



- Make sure file electrode is making good contact with the file.
- Touch the file with the contrary electrode and make sure the meter goes all the way to its end and there are no segments that do not light up.

- ! Watch out because the motor might start up when you do this.

⚠ WARNING

- Use only the special cable provided. Other cable could be electrically risky and result in damage or injury. Make sure the file goes all the way in. Give the file a light tug to make sure it is properly installed.
- Check the meter activity before each patient and do not use the unit if all the segments of the display do not light up. This suggests that the meter cannot make an accurate reading.

⚠ CAUTION

- Make sure the plugs go straight in.
- After insertion give plugs a light tug to make sure they are securely connected. Otherwise, data may not be transmitted accurately.
- Do not bump the plugs or drop anything on them when they are plugged in.

Meter Display

* For more information about canal measurement and usage precautions, refer to the user manual for the EndoSync A.I.



i. The Measurement Bar shows the location of the file tip. The Flash Bar blinks when the file is inside the canal.

ii. The 0.5 meter reading shows where the file tip is about 0.5 to 1.0 mm from the anatomical apex.

* The numbers 1, 2, and 3 on the meter do not indicate length in millimeters.

iii. If the file tip goes past the Flash Bar, an alarm will sound and the backlight will blink.



Operation

1

Turn on the EndoSync PLUS and EndoSync A.I. The backlight for the display will be yellow.

Hook the Contrary Electrode in the corner of the patient's mouth.



⚠ WARNING


- In some cases such as a blocked root canal, a measurement cannot be made. (For details refer to the section of the EndoSync A.I. manual that covers canals not suitable for measurement.)
- Accurate measurement is not always possible, especially in cases of abnormal or unusual root canal morphology; always take an X-ray to check the measurement results.
- If the meter does not move when the file is inserted, the unit may be malfunctioning and must not be used.
- Do not use an ultra sonic scaler while the contrary electrode is hooked in the patient's mouth; noise from the scaler could cause the motor to start running resulting in an accident or injury.
- Absolutely never allow the contrary electrode, the handpiece file electrode or the connections for these to contact an ordinary AC power source such as a socket; this could result in a very serious and dangerous shock.

⚠ CAUTION

- Occasionally the meter will make a sudden and large movement as soon as the file is inserted into the root canal, but it will return to normal as the file is advanced down towards the apex.
- The contrary electrode, file electrode and metal parts of the contra angle could cause an adverse reaction if the patient has an allergy to metals. Ask the patient about this before using the EndoSync PLUS.
- Take care that medicinal solutions such as formalin cresol (FC) or sodium hypochlorite do not get on the contrary electrode or the contra angle. These could cause an adverse reaction such as inflammation.
- The file electrode cannot be used with the following types of files. Use these files without attaching the file electrode. Files with a shank diameter greater than 1.2 mm, Files with shanks that do not have a circular cross section, Gates Glidden Drills, Tools with large cutting heads such as largo burs.

Operation

2




Plus and Minus Switches

Select a memory number (M1 to M6) with the Plus or Minus Switch.

* Before using motor handpiece, use a small hand file, such as #10 or #15, to penetrate the root canal manually down to the apical constriction.

* In some cases, a root canal cannot be measured because of an overflow of blood, saliva or chemicals or because the root canal is blocked.

3



The meter display appears when the file is inserted in to the canal. If the Auto Start and Stop is set to "ON", the motor will start running too.

* The numbers 1, 2, and 3 on the meter do not indicate length in millimeters but are used to estimate how far the file tip has gone down the canal.

* Press the Select Switch to change the display to the Torque display. Press it again to go back to the canal meter display.



Apical Action Function: ON

- When the measurement bar reaches the Flash Bar;
If the Apical Reverse or Stop is set to "Stp", the motor will stop and a continuous beep will sound.
If the Apical Reverse or Stop is set to "Rev", the motor will stop, a continuous beep will sound, and the motor will rotate in the non-cutting direction.
- When the measurement bar reaches the apex;
The motor will stop automatically and forcibly.

Apical Action Function: OFF

- When the measurement bar reaches the Flash Bar;
The motor will not stop and it will keep rotating in the cutting direction based on the memory setting.
- When the measurement bar reaches the apex;
The motor will not stop and it will keep rotating in the cutting direction based on the memory setting.

If the load on the file exceeds the reverse torque setting, the motor will stop and then reverse its rotation.*
A rapid, repeated three-toned beep will sound when this happens.

The motor will stop when the file is taken out of the canal.*
Gradually increase the size of the file until the root canal preparation is completed.

If necessary, prepare the apical seat.

(* Depends on setting.)

! If the canal is very dry, the Auto Start may not be triggered; in this case, press the Main Switch to start the motor.

WARNING

- Be sure to check that the memory setting for cutting direction and cutting direction of the file match. If they are not identical, the file could perforate the apical foramen when the Apical Reverse or Stop is set to "Rev".
- Accurate measurements cannot be made in some cases because of shape or other conditions. Always check the measurement with an X-ray.
- Do not let the file or metal parts of the contra angle touch the oral mucosa. This could cause the motor to start running and result in injuring the patient.
- An accurate measurement cannot be made if all the connectors are not properly plugged in. If the meter does not move along with the file, stop using the unit and check all the connections.


CAUTION

- The meter may not appear if the canal is infected or extremely dry. In this case, put a little hydrogen peroxide or saline solution in the canal but do not let it overflow.

Rotor Axle and Built-in Electrode Cleaning


* If the bars flicker during use, or if all the bars in the meter do not light up when the file touches the contrary electrode, clean the rotor axle and the built-in electrode in the following way.

1




Take out the screw and then take out the built-in electrode.

2



Put a little ethanol (70 vol% to 80 vol%) on a brush and clean the rotor axle with it.


3



Clean the built-in electrode with the brush.




4



Blow air on the electrode to remove any remaining moisture.


5



Hold the file release button and slide the guide bar straight in as shown in the illustration. Then rotate it left and right.



6




Slide the built-in electrode onto the guide bar and line up the screw holes.

CAUTION


- Do not bend or deform the electrode.
- Always use the guide bar and make sure it will not come out. If the guide bar is not properly fix in place, the internal contact could be bent, and then the unit might not be able to make accurate measurements or else it might malfunction.
- Do not run the motor with the guide bar inserted; this could damage the unit.

7



Slowly turn the screw and make sure the built-in electrode goes into the head properly.


8



Tighten the screw up securely and then hold down the file release button and pull out the guide bar.



9



Autoclave the contra angle.
 ☞ p. 25 "Autoclavable Components"

⚠ WARNING

- Make sure the screw is tight enough. Otherwise, it might come out and be swallowed. Also, measurements might not be accurate.

Maintenance

There are three ways to clean and disinfect components depending on the component. Be sure to follow the procedure below when performing daily maintenance.



Autoclavable Components



Contra Angle



Handpiece Rest

Wipe with Ethanol (70 vol% to 80 vol%)



Motor Handpiece



Power Supply Cord



Charger



Transmission Cable

Wash and Wipe with Ethanol (70 vol% to 80 vol%)



Guide Bar

Autoclavable Components

Maintenance Procedure

Cleaning

Disinfection

Lubrication

Packing

Sterilization

● Components maintained this way:

* Only the contra angle needs to be lubricated.




Contra Angle



Handpiece Rest

Cleaning



- (1) Disconnect the contra angle from the motor handpiece. Clean off the cutting debris with running water and a soft brush and then wipe off the water.

⚠ CAUTION


- Be careful to avoid cross contamination when performing maintenance.
- Before cleaning the contra angle, be sure to take out the file.
- If a medical agent being used for the treatment has adhered to the components, wash it off in running water.
- Do not clean the contra angle with an ultra sonic cleaning device.



(2) Use a three-way syringe etc. to blow out any moisture remaining inside the contra angle.



Disinfection



Wipe the components with a piece of gauze that has been dampened with ethanol (70 vol% to 80 vol%) and wrung out thoroughly.



Operating conditions for high-temperature washer-disinfectors

* When using a high-temperature washer-disinfector to clean the contra angle, strictly adhere to the conditions specified below.

High-temperature cleaning conditions

Unit Name	Mode	Detergent (concentration)	Neutralizer* (concentration)	Rinse (concentration)
Miele G7881	Vario TD	neodisher MediClean (0.3% - 0.5%)	neodisher Z (0.1% - 0.2%)	neodisher Mieleclear (0.02% - 0.04%)

* After cleaning there may be streaks or white spots on the contra angle. Use a neutralizer only if there are streaks or white spots.

Operating Precautions

- Always use a handpiece holder when washing the contra angle, making sure to rinse the inside of the contra angle thoroughly.
- If any medical agent remains inside the contra angle, it may corrode, resulting in a malfunction of the contra angle.
- For details on handling medical agents or adjusting their concentration, refer to the user manual for the washing device.
- Check to see if the contra angle including its inside, is completely dry. If any water remains inside the contra angle, expel it with an air gun etc. Failure to do so could result in the remaining water coming out during use and cause poor lubrication or sterilization.
- Always lubricate the contra angle after washing.




- ! Inappropriate cleaning methods and solutions will damage the contra angle.
- ! Do not clean the contra angle using strong acidic or alkaline solutions that could cause the metal to corrode.
- ! Do not leave the contra angle in the high-temperature washer-disinfector.

CAUTION

- Check to see if the contra angle including its inside, is completely dry. If any water remains inside the component, expel it with an air gun or another such tool. Failure to do so could result in the remaining water coming out during use and cause malfunction, or poor lubrication and sterilization.
- If dust or other impurities enter the contra angle, they may cause poor rotation.
- Do not use anything except ethanol (70 vol% to 80 vol%). Do not use too much ethanol as it could seep inside and damage the contra angle.
- Do not immerse the components in or wipe it with any of the following: functional water (acidic electrolyzed water, strong alkaline solution, or ozone water), medical agents (glutaral, etc.), or any other special types of water or commercial cleaning liquids. Such liquids may result in metal corrosion and adhesion of the residual medical agent to the components.
- Never clean the components with chemicals such as formalin cresol (FC) and sodium hypochlorite. These will damage the plastic parts of the components. If any of these liquids being applied to the components, wash it off in running water.
- Use only ethanol (70 vol% to 80 vol%) and OPTI-CIDE-3™ Surface Wipes for cleaning. Any other cleaning chemical or products should not be used including but not limited to the following cleaning products and similar cleaning products listed below because of the potential damage to the plastic components of the EndoSync PLUS.
 - CaviWipes™ • CaviCide™ • SANI-CLOTH™

Lubrication

* Only the contra angle needs to be lubricated.



Packing



Before autoclaving, the contra angle must be lubricated with the EndoSync Oil.

(1) Place the contra angle in a paper cup with the connection end facing up.



(2) Put 5 drops of the EndoSync Oil on the gear and wait for 10 minutes.



(3) Put a drop of the EndoSync Oil in each of the two points between the built-in electrode and the head as indicated by the arrows in the illustration.

(4) Take the contra angle out of the paper cup and wipe off any excess oil which may have seeped out. Dampen a piece of gauze with ethanol (70 vol% to 80 vol%), wring it out and then wipe the contra angle with it.

! Do not use anything except ethanol (70 vol% to 80 vol%) for cleaning. Never wipe the contra angle with solutions containing formalin cresol (FC) or sodium hypochlorite, which damage plastic; wipe them off immediately if they accidentally get on the contra angle.

! Do not immerse in any fluid.

! Do not connect the contra angle to the motor handpiece immediately after lubrication for use or charging. Otherwise the oil seep inside the motor handpiece and it might malfunction.

Put components in individual autoclave pouches.
For U.S. customers, use only FDA-cleared pouches.

CAUTION

- Do not use any type of spray other than the EndoSync Oil.
- Failure to lubricate the contra angle will result in a malfunction.
- Put the cap on after use. Oil could seep out if the container is tipped over or the nozzle points down.
- After lubricating, wipe oil from the outside of the nozzle. Otherwise oil may seep out from under the cap.
- Leave the contra angle in the paper cup for at least 10 minutes so that the oil is thoroughly absorbed by the contra angle mechanism.

Sterilization

Recommended Temperature and Time

Sterilizer Type: Gravity

Temperature	Time	Drying time after sterilization
+132 °C (+269.6 °F)	15 minutes	15 minutes
+121 °C (+249.8 °F)	30 minutes	15 minutes

(For U.S. customers)

Temperature	Time	Drying time after sterilization
+134 °C (+273.2 °F)	min. 6 minutes	min. 10 minutes
+121 °C (+249.8 °F)	min. 60 minutes	min. 10 minutes

Sterilizer Type: Dynamic Air Removal

Temperature	Time	Drying time after sterilization
+134 °C (+273.2 °F)	3 minutes	10 minutes

Autoclave the components.



- ! Thoroughly clean and wash the components before autoclaving. If chemical solutions or foreign debris are not removed, autoclaving could damage or deform the components.
- ! The sterilization and drying temperatures must not exceed +135°C (+275°F).
- ! No components can be autoclaved other than the contra angle, hand-piece rest.
- ! Take the file out of the contra angle before autoclaving.
- ! Be sure to lubricate the contra angle with the EndoSync Oil before autoclaving it.
- ! Follow file manufacturer's recommendations for autoclaving files.

Wipe with Ethanol (70 vol% to 80 vol%)

Maintenance Procedure

Disinfection

- Components maintained this way:



Motor Handpiece



Charger




Power Supply Cord



Transmission Cable

Disinfection



Wipe the components with a piece of gauze that has been dampened with ethanol (70 vol% to 80 vol%) and wrung out thoroughly.



! WARNING

- To prevent the spread of serious, life-threatening infections such as HIV and hepatitis B, the components (contra angle, handpiece rest) must be autoclaved after each patient's treatment has been completed.

! CAUTION

- Do not sterilize the components by any method other than autoclaving.
- Components are extremely hot right after autoclaving. Wait for them to cool off before touching.
- Do not leave the components in the autoclave.
- Do not use anything except ethanol (70 vol% to 80 vol%). Do not use too much ethanol as it could seep inside and damage the components. Do not apply or spray with any fluid.
- Do not immerse the components in or wipe it with any of the following: functional water (acidic electrolyzed water, strong alkaline solution, or ozone water), medical agents (glutaral, etc.), or any other special types of water or commercial cleaning liquids. Such liquids may result in metal corrosion and adhesion of the residual medical agent to the components.
- Never clean the components with chemicals such as formalin cresol (FC) and sodium hypochlorite. These will damage the plastic parts of the components. If any of these liquids being applied to the components, use dry gauze etc. to wipe it off.
- Use only ethanol(70 vol% to 80 vol%) and OPTI-CIDE-3™ Surface Wipes for cleaning. Any other cleaning chemical or products should not be used including but not limited to the following cleaning products and similar cleaning products listed below because of the potential damage to the plastic components of the EndoSync PLUS.
 - CaviWipes™
 - CaviCide™
 - SANI-CLOTH™

Wash and Wipe with Ethanol (70 vol% to 80 vol%)

Maintenance Procedure

Cleaning → Disinfection

- Components maintained this way:



Guide Bar


Cleaning



Clean off the cutting debris in running water with a soft brush and then wipe off the water.



Disinfection



Wipe the component with a piece of gauze that has been dampened with ethanol (70 vol% to 80 vol%) and wrung out thoroughly.



⚠ CAUTION

- Do not clean the component with an ultra sonic cleaning device.
- Do not use anything except ethanol (70 vol% to 80 vol%).
- Do not immerse the component in or wipe it with any of the following: functional water (acidic electrolyzed water, strong alkaline solution, or ozone water), medical agents (glutaral, etc.), or any other special types of water or commercial cleaning liquids. Such liquids may result in metal corrosion and adhesion of the residual medical agent to the component.
- Never clean the component with chemicals such as formalin cresol (FC) and sodium hypochlorite. These will damage the plastic parts of the components. If any of these liquids being applied to the component, wash it off in running water.
- Use only ethanol (70 vol% to 80 vol%) and OPTI-CIDE-3™ Surface Wipes for cleaning. Any other cleaning chemical or products should not be used including but not limited to the following cleaning products and similar cleaning products listed below because of the potential damage to the plastic component of the EndoSync PLUS.
 - CaviWipes™
 - CaviCide™
 - SANI-CLOTH™

Replacement Parts and Storage

(1) Replacement Parts

* Replace the parts as necessary depending on degree of wear and length of use.


* Order parts from Brasseler USA.

Battery Replacement

Replace the battery when it starts to lose power relatively quickly after being fully charged. The battery will last for approximately 1 year under normal circumstances and use.



1



Turn power off. Remove the screw and slide the battery cover off.


2



Take out the old battery and disconnect it.

! Do not leave the power on when disconnecting the battery.

3



a) Remove an old cushion.
b) Peel off the paper shield from the sticky side of the cushion and stick it to the bottom of the main unit.

! A cushion must be pasted on to fill a gap between the battery and the main unit.

4



Connect the battery cord and then slide the battery along the bottom.

! Dispose of old lithium ion batteries in an environmentally safe way and in strict accordance with local regulations.

5



Replace the cover and its screw.

! Be careful not to pinch the battery cord when replacing the cover.
! Do not tighten the cover screw too much; this could strip the threads.


CAUTION

- Use only the battery designed for the EndoSync PLUS. Other types could cause overheating.
- Do not use a battery if it is leaky, deformed, discolored or if its label is peeled off. It might overheat.

Built-in Electrode Replacement




1




Take out the screw and then take out the built-in electrode.

2




Put a little ethanol (70 vol% to 80 vol%) on a brush and clean the rotor axle with it.

3




Blow air on the electrode to remove any remaining moisture.

4




Hold the file release button and slide the guide bar straight in as shown in the illustration. Then rotate it left and right. 

5



Slide the built-in electrode onto the guide bar and line up the screw holes.

6




Slowly turn the screw and make sure the built-in electrode goes into the head properly.

CAUTION

- Always use the guide bar and make sure it will not come out. If the guide bar is not properly fix in place, the internal contact could be bent, and then the unit might not be able to make accurate measurements or else it might malfunction.
- Do not run the motor with the guide bar inserted; this could damage the unit.


7



Tighten the screw up securely and then hold down the file release button and pull out the guide bar.



8



Autoclave the contra angle.
☞ p. 25 "Autoclavable Components"

⚠ WARNING

• Make sure the screw is tight enough. Otherwise, it might come out and be swallowed. Also, measurements might not be accurate.

(2) Transport and Storage Environments

Temperature: -10°C to +45°C (+14°F to +113°F)

Humidity: 10% to 85 % (without condensation)

Atmospheric Pressure: 70 kPa to 106 kPa

- * Do not expose to direct sunlight frequently or for long times.
- * If the unit has not been used for a long time, make sure it works properly before using.
- * Always remove the battery prior to storing or shipping the unit.

Inspection and Warranty

Regular Inspection


- * Maintenance and inspection are generally considered to be the duty and obligation of the user, but if, for some reason, the user is unable to carry out these duties, contact Brasseler USA for technical support.
- * Replace the parts listed in the Parts Lists as necessary depending on degree of wear and length of use.
- * This apparatus should be inspected every 6 months in accordance with the following maintenance and inspection items.

Inspection Items

1. Check that the battery does not seem to be losing its charge too quickly.
2. Check that pressing the Main Switch turns the unit on. After the unit is on, check that pressing the Main Switch turns the motor on and off. Check that the unit turns off when the Main Switch is pressed while the Select Switch is being held down.
3. Check that pressing the Plus and Minus Switches changes the memory number from M1 through M6.
4. Check that the settings for each memory can be changed.
5. Make sure the connection end of the motor handpiece is not damaged or dirty.
6. Make sure that the connection end of the contra angle is not damaged or dirty and that it can be securely connected to the motor handpiece. Make sure that the file release button operates properly and that files can be securely installed.
7. When used with the EndoSync A.I., touch the file with the contrary electrode and make sure that all the segments for the meter light up properly.

* For repairs contact Brasseler USA.

Parts List

EndoSync Head W/ Built-in Electrode (Contra Angle) REF No. 5025561U0	EndoSync PLUS Motor Handpiece (Motor Handpiece) REF No. 5027418U0	Built-in Electrode W/ Guide Bar (Built-in Electrode with Guide Bar) REF No. 5025560U0	EndoSync Oil 60 ml REF No. 5025559U0
			
Lithium-Ion Battery (Battery) REF No. 5026034U0	Handpiece Rest REF No. 5025566U0	EndoSync Charger (Charger) REF No. 5025562U0	Power Supply Cord REF No. 5025564U0
			
Transmission Cable REF No. 5025552U0			
			

Maintenance and Inspection Items

■ Disposal of Medical Devices

Any medical devices which could possibly be contaminated must be first decontaminated by the responsible doctor or medical institution and then be disposed of in accordance with local laws and regulations.

The rechargeable battery should be recycled. Metal parts of the equipment are disposed as scrap metal. Synthetic materials, electrical components, and printed circuit boards are disposed as electrical scrap. Material must be disposed according to the relevant national legal regulations. Consult specialized disposal companies for this purpose. Please inquire of the local city/community administrations concerning local disposal companies.

■ Service

The EndoSync PLUS should be repaired and serviced by Brasseler USA and Brasseler Canada technicians.

- For customers in the U.S., call 1-800-841-4522.
- For customers in Canada, call 1-800-363-3838.

Warranty

■ 1 Year Limited Warranty

1. Brasseler USA gives a guarantee for one year beginning from the date of purchase. Within this period any defect that is due to faulty manufacturing or material will be remedied by repair or replacement at the judgment of Brasseler USA.
2. Warranty repair and service: In the event of a claim under this guarantee, the device is to be sent to Brasseler USA. For customers in the U.S., call 1-800-841-4522. For customers in Canada, call 1-800-363-3838.
3. In the case of damage caused by wear and tear, careless handling and repairs not carried out by Brasseler USA, the warranty ceases to be valid. This guarantee may not form the basis for any claims for damages, in particular not for compensation of consequential damages.

The buyer assumes responsibility for damage due to dropping of the unit, improper use and utilization of product and chemicals other than those stated in this instruction manual for cleaning. It is the customer's responsibility to maintain the exact rated voltage indicated at the bottom of the unit, and the office maintains electrical outlets for proper performance of the charger.

4. This warranty does not include the external accessories, built-in electrode or batteries.

Troubleshooting

If the unit does not seem to be working properly, the user should first try to inspect and adjust it himself.

* If the user is unable to inspect the unit himself or if the unit fails to work properly after being adjusted or after parts are replaced, contact Brasseler USA.

Problem	Check Points	Response
Does not turn on.	<ul style="list-style-type: none"> • Check battery power. • Check battery installation. 	<ul style="list-style-type: none"> • Charge battery • Install battery properly.
No beeping sound.	<ul style="list-style-type: none"> • Check if the Beeper Volume is set to "OFF". 	<ul style="list-style-type: none"> • Set it to "Low" or "High".
Beep sounds even when unit is not being used.	<ul style="list-style-type: none"> • Unit may be set to Normal Mode (CCW). 	<ul style="list-style-type: none"> • A beep sounds periodically whenever the unit is set to Normal Mode (CCW). Set the beeper volume to "OFF" if it is annoying. (This will stop all beeping except when the unit is turned on.)
Backlight color does not change.	<ul style="list-style-type: none"> • Check if the Backlight Color Change is set to "OFF". 	<ul style="list-style-type: none"> • Set it to "ON", if necessary.
Motor does not start when file is inside canal.	<ul style="list-style-type: none"> • Is the EndoSync A.I. properly connected and turned on? • Is the contrary electrode for the EndoSync A.I. hooked in the patient's mouth? • Is the Apical Action Function set to "OFF"? • Is Auto Start & Stop set to "OFF"? • Has the meter gone past the Flash Bar? 	<ul style="list-style-type: none"> • Check transmission cable connections. Then turn on the EndoSync A.I. • Hook the contrary electrode in the corner of the patient's mouth. • Set it to "ON". • Set it to "ON". • Set the Apical Reverse or Stop to "Rev" (reverse).
File does not advance smoothly.	<ul style="list-style-type: none"> • Is the cutting direction of the file correct? • Is the cutting direction setting appropriate for the inserted file? 	<ul style="list-style-type: none"> • Check that the correct file is inserted. • Check the setting for the Cutting Direction of File and match with the file.
Motor starts but then stops right away.	<ul style="list-style-type: none"> • Did you hold down the Main Switch for more than 1 second? • Does "Abn.Stop LowBat" appear in the display? 	<ul style="list-style-type: none"> • If you hold the Main Switch down for more than 1 second, the motor runs only while the switch is held down and stops when it is released. The motor will run without stopping if you release the switch in less than 1 second. • Very low battery power. Charge battery.
Motor reverses rotation on its own.	<ul style="list-style-type: none"> • Check the Reverse Torque setting. • Check if the Apical Reverse or Stop is set to "Rev" (reverse). • Is the Cutting Direction of File set to "CCW"? 	<ul style="list-style-type: none"> • The Torque Reverse can be turned off with "TRL" setting. • You can change it to "Stp" (stop). • Set it to "CW".
Motor reverses rotation too quickly.	<ul style="list-style-type: none"> • Check the Reverse Torque setting. • Is the Apical Torque Reduction set to "ON"? 	<ul style="list-style-type: none"> • Increase its setting value. • The torque reverse value goes down as the file approaches the apex if the Apical Torque Reduction is set to "ON". Set it to "OFF" to keep the torque reverse value constant.
Motor runs back and forth continuously	<ul style="list-style-type: none"> • Is Rotation Mode set to "OTR"? • Does it do this even after calibration? 	<ul style="list-style-type: none"> • Torque load is greater than the setting for OTR Mode. • Increase the setting for the Reverse Torque by 1 level.  p. 14 "Reverse Torque Settings:"

Problem	Check Points	Response
Meter is not stable during use.	<ul style="list-style-type: none"> • Does the built-in electrode need replacement? Has it been replaced recently? • Is the screw for the built-in electrode loose? 	<ul style="list-style-type: none"> • Clean and lubricate the contra angle. • Take out the built-in electrode and clean it and the rotor axle with a brush. • Replace the built-in electrode. • Tighten the screw.
Motor handpiece will not go in reverse rotation.	<ul style="list-style-type: none"> • Is the Reverse Torque set to “TRL” (torque reverse less)? • Is Reverse Torque setting too high? • Is the Apical Action Function set to “OFF”? • Is the Apical Reverse or Stop set to “Stp” (stop)? 	<ul style="list-style-type: none"> • Change the setting to other than “TRL”. • Reduce the torque reverse value • Set it to “ON”. • Set it to “Rev” (reverse).
Micromotor changes speed on its own.	<ul style="list-style-type: none"> • Is the Apical Slow Down set to “ON”? • Is the Torque Slow Down set to “ON”? 	<ul style="list-style-type: none"> • When this is set to “ON”, the motor slows down as the file approaches the apex.  p. 15 “Apical Slow Down**.” • When this is set to “ON”, the motor slows down as the torque increases.  p. 15 “Torque Slow Down.”
Unit turns off/on its own.	<ul style="list-style-type: none"> • Was the unit no used for a long time? • Does “Please Charge” appear in the display? • This can happen if the battery is very low and a large load is applied to the file. 	<ul style="list-style-type: none"> • Auto power off was probably activated. Press the Main Switch to turn the unit back on. • Battery must be charged right away. • Battery must be charged right away.
Cannot change the memory settings.	<ul style="list-style-type: none"> • Is the Memory Setting Lock set to “ON”? • Did you hold down the Select Switch? 	<ul style="list-style-type: none"> • Set it to “OFF” and unlock the setting function.  p. 17 “Memory Setting Lock.” • It is necessary to hold it for more than 2 seconds.
Error 01	<ul style="list-style-type: none"> • Turn the unit off and disconnect the transmission cable; does the same error message appear when the unit is turned back on? • Is there some debris on the connector for the transmission cable? • Does the error message appear when the transmission cable is twisted or bent sharply? 	<ul style="list-style-type: none"> • If the unit’s operation is restored by disconnecting the transmission cable, the problem was only temporary and there is nothing wrong with it. • If the same error occurs after disconnecting the transmission cable, there is probably something wrong with the unit. Contact Brasseler USA. • If there is, clean the connector. • There could be a broken wire inside the cable; replace it with a new one.
Error 04	<ul style="list-style-type: none"> • Does this happen repeatedly? 	<ul style="list-style-type: none"> • There may be something wrong with the control board. (In this case, memory settings cannot be saved, but they can still be changed even though they will not be saved.)
Error 06	<ul style="list-style-type: none"> • Does this happen repeatedly? 	<ul style="list-style-type: none"> • The motor circuits may be malfunctioning. Have the unit repaired.

Technical Specifications

Technical Specifications

Specifications may be changed without notice due to improvements.

Name	EndoSync PLUS
Model	TR-CM
Type	BSL PLUS
Degree of Protection (IEC 60529)	IPX0
Indications for Use	The EndoSync PLUS motorized handpiece can be used to enlarge and prepare root canals, remove gutta-percha points, and for professional tooth cleaning. When connected to the Brasseler USA EndoSync A.I. or J. MORITA ROOT ZX mini, an apex locator (sold separately), the EndoSync PLUS can be used to measure the length of root canals.
Operating Principle	By electric drive, it transmits motion, such as rotation, to treatment instruments (dental files, reamers, etc.).
Essential Performance	None (There is no unacceptable risk.)

Handpiece

Free Running Operation Speed	50 ±5 to 1000 ±100 r/min
Gear Ratio	1.9 : 1
Usable Burs	Type 1 (CA)
Rated Torque	Min. 0.04 Nm
Chuck Type	Push button latch type
Protection against Electric Shock	Internal powered ME equipment / Type BF applied part
Battery	Lithium ion battery (DC 3.7 V)
Dimensions	Max. diameter 28 ±3 mm × length 150 ±10 mm
Weight	Approx. 80 g
Applied Part	Contra angle, Motor handpiece

Battery Charger

Rated Input Voltage	AC 100 to 240 V, 50/60 Hz
Power Consumption	19 VA
Dimensions	68 × 108 × 85 mm
Weight	350 g (includes power supply cord)

Symbols

* Some symbols may not be used.



Serial Number

Rx Only

CAUTION:
Federal law restricts this device to sale by
or on the order of a dentist. (Valid only for
U.S.A.)



Date of manufacture




GS1 DataMatrix



cTUVus certification mark
(Valid only for U.S.A. and Canada)



Keep away from rain



Class II Equipment



This way up



Type BF applied part




Fragile



Refer to instructions for use




Temperature limitation



Autoclavable up to 135°C



Humidity limitation



Atmospheric pressure limitation

Electromagnetic Disturbances (EMD)

The EndoSync PLUS (Model: TR-CM, hereafter “this device”) conforms to IEC 60601-1-2:2014 Ed. 4.0, the relevant international standard for electromagnetic disturbances (EMD).

The following is the “Guidance and Manufacturer’s Declaration” which is required by IEC 60601-1-2:2014 Ed. 4.0, the relevant international standard for electromagnetic disturbances.

This is a Group 1, Class B product according to EN 55011 (CISPR 11).

This means that this device does not generate and/or use internationally radio-frequency energy, in the form of electromagnetic radiation, inductive and/or capacitive coupling, for the treatment of material or inspection/analysis purpose and that it is suitable for use in domestic establishments and in establishments directly connected to a low voltage power supply network which supplies buildings use for domestic purposes.

Guidance and Manufacturer’s Declaration – Electromagnetic Emissions		
This device is intended for use in the electromagnetic environment specified below. The customer or the user of this device should assure that it is used in such an environment.		
Emissions Test	Compliance	Electromagnetic Environment – Guidance
Conducted disturbance CISPR 11	Group 1 Class B	This device uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. This device is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Radiated disturbance CISPR 11	Group 1 Class B	
Harmonic current ¹ IEC 61000-3-2	Class A	
Voltage fluctuations and flicker IEC 61000-3-3	Clause 5	

*1: Although this device is not applicable to Harmonics test since the rated power is less than 75 W, it has been tested as a reference according to limits for Class A.

WARNING

- The use environment of this device is the Home healthcare environment.
- This device needs special precautions regarding EMD and needs to be installed and put into service according to the EMD information provided in the AC-COMPANYING DOCUMENTS.
- Use of parts other than those accompanied or specified by J. MORITA MFG. CORP. could result in increased electromagnetic emissions or decreased electromagnetic immunity of this device and result in improper operation.
- Do not use this device as adjacent or stacked as possible with other. When adjoining or stacking is necessary, use it after observing whether this equipment and other equipment work properly.
- Portable and mobile RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm to any part of the TR-CM, including cables specified by the manufacturer.

Guidance and Manufacturer’s Declaration – Electromagnetic Immunity			
This device is intended for use in the electromagnetic environment specified below. The customer or the user of this device should assure that it is used in such an environment.			
Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment – Guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	±8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transients/bursts IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	±2 kV for power supply lines ¹ ±1 kV for input/output lines ¹	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	<u>AC/DC power</u> ±0.5 kV, ±1 kV line(s) to line(s) ±0.5 kV, ±1 kV, ±2 kV line(s) to earth <u>Signal input/output</u> ±2 kV line(s) to earth	<u>AC/DC power</u> ±0.5 kV, ±1 kV line(s) to line(s) ±0.5 kV, ±1 kV, ±2 kV line(s) to earth <u>Signal input/output</u> ² ±2 kV line(s) to earth	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply lines IEC 61000-4-11	<u>dips</u> 0% U_T : 0.5 cycle (at 0, 45, 90, 135, 180, 225, 270, 315°) 0% U_T : 1 cycle (at 0°) 70% U_T : 25/30 cycles (at 0°) 25 (50 Hz)/30 (60 Hz) <u>short interruptions</u> 0% U_T : 250/300 cycles 250 (50 Hz)/300 (60 Hz)	<u>dips</u> 0% U_T : 0.5 cycle (at 0, 45, 90, 135, 180, 225, 270, 315°) 0% U_T : 1 cycle (at 0°) 70% U_T : 25/30 cycles (at 0°) 25 (50 Hz)/30 (60 Hz) <u>short interruptions</u> 0% U_T : 250/300 cycles 250 (50 Hz)/300 (60 Hz)	Mains power quality should be that of a typical commercial or hospital environment. If user of this device requires continued operation during power mains interruptions, it is recommended that this device be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	30 A/m (r.m.s.) 50 Hz or 60 Hz	30 A/m (r.m.s.) 50 Hz or 60 Hz	Power frequency magnetic field should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE 1: U_T is the a.c. mains voltage prior to application of the test level. NOTE 2: r.m.s.: root mean square			

*1: This test is not applicable since the EUT signal cable is less than 3 m.

*2: Not applicable because it does not connect directly to outdoor cable.

Guidance and Manufacturer's Declaration – Electromagnetic Immunity			
This device is intended for use in the electromagnetic environment specified below. The customer or the user of this device should assure that it is used in such an environment.			
Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment – Guidance
Conducted RF IEC 61000-4-6	3 V ISM ^(c) / amateur radio frequency band: 6 V 150 kHz to 80 MHz	3 V ISM ^(c) / amateur radio frequency band: 6 V 150 kHz to 80 MHz	Portable and mobile RF communications equipment should be used no closer to any part of this device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Radiated RF IEC 61000-4-3	10 V/m 80 MHz to 2.7 GHz 27 V/m 385 MHz 28 V/m 450 MHz 9 V/m 710, 745, 780 MHz 28 V/m 810, 870, 930, MHz 28 V/m 1720, 1845, 1970 MHz 28 V/m 2450 MHz 9 V/m 5240, 5500, 5785 MHz	10 V/m 80 MHz to 2.7 GHz 27 V/m 385 MHz 28 V/m 450 MHz 9 V/m 710, 745, 780 MHz 28 V/m 810, 870, 930, MHz 28 V/m 1720, 1845, 1970 MHz 28 V/m 2450 MHz 9 V/m 5240, 5500, 5785 MHz	Recommended separation distances $d = 1.2 \sqrt{P}$ 150 kHz to 80 MHz $d = 0.4 \sqrt{P}$ 80 MHz to 800 MHz $d = 0.7 \sqrt{P}$ 800 MHz to 2.7 GHz $d = \frac{6}{E} \sqrt{P}$ Portable wireless RF communication equipment Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer, E is the compliance level in V/m and d is the recommended separation distance in meters (m). Field strengths from field RF transmitters, as determined by an electromagnetic site survey ^(a) , should be less than the compliance level in each frequency range ^(b) . Interference may occur in the vicinity of equipment marked with the following symbol: 
NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies. NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
(a) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which this device is used exceeds the applicable RF compliance level above, this device should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating this device. (b) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m. (c) The ISM (Industrial, Scientific and Medical) bands between 0.15 MHz and 80 MHz are 6.765 MHz to 6.795 MHz; 13.553 MHz to 13.567 MHz; 26.957 MHz to 27.283 MHz; and 40.66 MHz to 40.70 MHz. The amateur radio bands between 0.15 MHz and 80 MHz are 1.8 MHz to 2.0 MHz, 3.5 MHz to 4.0 MHz, 5.3 MHz to 5.4 MHz, 7 MHz to 7.3 MHz, 10.1 MHz to 10.15 MHz, 14 MHz to 14.2 MHz, 18.07 MHz to 18.17 MHz, 21.0 MHz to 21.4 MHz, 24.89 MHz to 24.99 MHz, 28.0 MHz to 29.7 MHz and 50.0 MHz to 54.0 MHz.			

Essential Performance

None

Cable List

No.	Interface(s):	Max. Cable Length, Shielding	Cable Classification
1.	AC Power Cable	1.5 m, Un-shielded	AC Power Line
2.	Probe Cord	1.7 m, Un-shielded	Signal Line (Patient-Coupled Cable)
3.	Communication Cable	1.6 m, Un-shielded	Signal Line



BRASSELER
USA®

One Brasseler Boulevard, Savannah, GA 31419, U.S.A.

